

## Information on EU Directive 2014/68/EU On pressure equipment

**Manufacturer:** VSE Volumentchnik GmbH, 58809 Neuenrade

In terms of Article 2, No. 5 of the directive named above, VSE volume sensors are so-called „pressure-maintaining components“ and this directive thereby relates to them.

VSE volume sensors must thereby comply with the technical requirements named in Section 4 of the directive in accordance with Article 4, Paragraph (1) d, Piping according to Paragraph (1) c.

Typically, the fluids measured fall into Group 2 in accordance with Article 13, Paragraph 1, b. The volume sensors sold by VSE do not comply with the limit values defined under Article 4, Paragraph 1, a.

The technical requirements on volume sensors from VSE are therefore limited to the criteria defined in Article 4, Paragraph 3. That means that the devices must be designed and manufactured in accordance with the good engineering practices prevailing in the member state. We hereby confirm this. The paragraph also states that these units may not bear the CE label named in Article 18.

A CE declaration of conformity is therefore not issued in accordance with 2014/68/EU.  
The CE label of our volume sensors refers to Directive 2014/30/EU.

Neuenrade, April 27th, 2016

VSE Volumentchnik GmbH



Axel Vedder  
Managing Director